

(iii) All proficiency testing reports received are reviewed by the appropriate staff to evaluate the laboratory's performance and to identify any problems that require corrective action; and

(iv) An approved corrective action plan is followed when any proficiency testing results are found to be unacceptable or unsatisfactory;

(5) Ensure that the quality control and quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur;

(6) Ensure the establishment and maintenance of acceptable levels of analytical performance for each test system;

(7) Ensure that all necessary remedial actions are taken and documented whenever significant deviations from the laboratory's established performance specifications are identified, and that patient test results are reported only when the system is functioning properly;

(8) Ensure that reports of test results include pertinent information required for interpretation;

(9) Ensure that consultation is available to the laboratory's clients on matters relating to the quality of the test results reported and their interpretation concerning specific patient conditions;

(10) Employ a sufficient number of laboratory personnel with the appropriate education and either experience or training to provide appropriate consultation, properly supervise and accurately perform tests and report test results in accordance with the personnel responsibilities described in this subpart;

(11) Ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results;

(12) Ensure that policies and procedures are established for monitoring individuals who conduct preanalytical, analytical, and postanalytical phases

of testing to assure that they are competent and maintain their competency to process specimens, perform test procedures and report test results promptly and proficiently, and whenever necessary, identify needs for remedial training or continuing education to improve skills;

(13) Ensure that an approved procedure manual is available to all personnel responsible for any aspect of the testing process; and

(14) Specify, in writing, the responsibilities and duties of each consultant and each person, engaged in the performance of the preanalytic, analytic, and postanalytic phases of testing, that identifies which examinations and procedures each individual is authorized to perform, whether supervision is required for specimen processing, test performance or results reporting, and whether consultant or director review is required prior to reporting patient test results.

[57 FR 7172, Feb. 28, 1992, as amended at 68 FR 3713, Jan. 24, 2003]

**§ 493.1409 Condition: Laboratories performing moderate complexity testing; technical consultant.**

The laboratory must have a technical consultant who meets the qualification requirements of §493.1411 of this subpart and provides technical oversight in accordance with §493.1413 of this subpart.

**§ 493.1411 Standard; Technical consultant qualifications.**

The laboratory must employ one or more individuals who are qualified by education and either training or experience to provide technical consultation for each of the specialties and subspecialties of service in which the laboratory performs moderate complexity tests or procedures. The director of a laboratory performing moderate complexity testing may function as the technical consultant provided he or she meets the qualifications specified in this section.

(a) The technical consultant must possess a current license issued by the State in which the laboratory is located, if such licensing is required.

(b) The technical consultant must—